



Günter Keul GmbH
Von-Langen-Weg 10
D-48565 Steinfurt
Tel.: 02551 / 2097 Fax.: 02551 / 80883

Micr-o-trol CRP



CONTROL

ASSAY VALUES AND EXPECTED RANGES
Wertebblatt und Wertebereich

CBC-3D+CRP



B0825C

2025-11-05

Instruments :											
Instrument	Parameter / Paramètre		CONTROL		L	CONTROL		N	CONTROL		H
			LOT		B0825CL	LOT		B0825CN	LOT		B0825CH
			Mean	Limit	Mean	Limit	Mean	Limit			
			Zielwert	Bereich	Zielwert	Bereich	Zielwert	Bereich			
MICROS CRP Series	WBC/Leuko	10 ³ /μL & 10 ⁹ /L	2.0	0,4	7.9	±	1,0	20.2	2,2		
	RBC/Ery	10 ⁶ /μL & 10 ¹² /L	2.25	0,15	4.58	±	0,18	5.60	0,25		
	Hgb	g/dL	6.2	0,4	13.6	±	0,5	17.8	0,7		
	Hct	%	16.8	2,0	38.2	±	2,5	50.4	3,0		
MICROSEMI LC-667	MCV/VGM	fL	74.5	5,0	83.5	±	5,0	90.0	5,0		
	RDW/IDR	%	14.9	5,0	14.4	±	5,0	14.4	5,0		
ZYBIO Z3 CRP	Plt/Thrombo	10 ³ /μL & 10 ⁹ /L	77	20	247	±	40	500	70		
	MPV/VPM	fL	7.9	2,0	7.5	±	2,0	7.7	2,0		
	LYM%	%	62.4	10.0	32.9	±	8.0	15.0	6.0		
	MID%	%	6.8	6.8	6.0	±	5.0	3.5	45780,0		
	GRA%	%	30.8	8.0	61.1	±	8,0	81.5	7,0		
	CRP	mg/dL	0.40	0,30	2.30	±	1,20	4.65	2,20		



Günter Keul GmbH
Von-Langen-Weg 10 D-48565 Steinfurt
Tel.: 02551 / 2097 Fax.: 02551 / 80883
Web: www.keul.de E-Mail: info@keul.de

