



ASSAY VALUES AND EXPECTED RANGES
SOLLWERTE UND BEREICHE

CONTROL

VALEURS CIBLES ET INTERVALLES DE VARIATIONS
RiliBÄK

LOT

B0825

QC-Monate

August, September, Oktober



2025-11-05

Geräte Instruments	Parameter	Einheit Unit	LOT B0825L			LOT B0825N			LOT B0825H		
Keul Hem-o-test 3600	WBC	×10 ⁹ /L	1,90	±	0,60	8,10	±	0,53	22,10	±	1,44
	Lym%	%	61,4	±	9	30,3	±	8,00	13,7	±	8
	Gran%	%	29,3	±	9	62,8	±	8,00	82,2	±	8
	Mid%	%	9,3	±	9,3	6,9	±	6,00	4,1	±	4,0
	Lym#	×10 ⁹ /L	1,16	±	0,3	2,45	±	0,70	3,02	±	2
	Gran#	×10 ⁹ /L	0,6	±	0,3	5,1	±	0,7	18,2	±	2
	Mid#	×10 ⁹ /L	0,18	±	0,18	0,56	±	0,5	0,91	±	0,8
	RBC	×10 ¹² /L	2,21	±	0,09	4,65	±	0,19	5,66	±	0,23
	HGB	g/L	57	±	2,28	136	±	5,44	184	±	7,36
	HGB	g/dl	5,7	±	0,228	13,6	±	0,544	18,4	±	0,736
	HCT	%	18,1	±	0,905	42,7	±	2,14	55,5	±	2,775
	MCV	fL	82,1	±	5	91,9	±	5	98	±	5
	MCH	pg	25,6	±	2,5	29,2	±	2,5	32	±	2,5
	MCHC	g/L	315	±	30	319	±	30	330	±	30
	RDW-CV	%	17,9	±	3	17,6	±	3	18,6	±	3
	RDW-SD	fL	50,1	±	8	55,4	±	8	61,6	±	8
	PLT	×10 ⁹ /L	69	±	9	266	±	23	539	±	40
	MPV	fL	8,9	±	3	8,5	±	3	8,7	±	3
	PDW	fL	11	±	3	10,3	±	3	10,8	±	3
	PCT	%	0,062	±	0,062	0,227	±	0,2	0,476	±	0,2
	P-LCR	%	20,1	±	10	17,3	±	10	19,7	±	10
	P-LCC	×10 ⁹ /L	14	±	14	47	±	25	108	±	30
	PDW	/	15,9	±	3	16	±	3	15,9	±	3



Bio-technie® - 19 Rue Louis Delourmel
35230 - NOYAL CHATILLON / SEICHE - FRANCE



R&D Systems, Inc. - 614 Mc Kinley Place N.E. - Minneapolis, MN USA 55413