



CBC-3D[®]

for Melet Schloesing & Dialab

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

 QCP DATA MONTHS: **AUGUST, SEPTEMBER, OCTOBER**
LOT
B0825

2025-11-05

Instrument	Parameter	Low		Normal		High	
		LOT	B0825L	LOT	B0825N	LOT	B0825H
Melet Schloesing MS4s	WBC × 10 ³ /μL		1.6 ± 0.5		7.2 ± 1.0		19.6 ± 2.5
Dialab DLB5	RBC × 10 ⁶ /μL		2.25 ± 0.22		4.65 ± 0.28		5.73 ± 0.34
	HGB g/dL		5.7 ± 0.6		13.6 ± 0.8		18.3 ± 1.1
	HGB g/L		57 ± 6		136 ± 8		183 ± 11
	HGB mmol/L		3.5 ± 0.4		8.4 ± 0.5		11.3 ± 0.8
	HCT %		16.4 ± 2.0		37.7 ± 2.4		50.7 ± 3.0
	HCT L/L		0.164 ± 0.020		0.377 ± 0.024		0.507 ± 0.030
	MCV fL		73.0 ± 5.0		81.0 ± 5.0		88.5 ± 5.0
	MCH pg		25.3 ± 2.5		29.2 ± 2.5		31.9 ± 2.8
	MCH fmol		1.57 ± 0.16		1.81 ± 0.17		1.98 ± 0.19
	MCHC g/dL		34.7 ± 3.0		36.1 ± 3.0		36.1 ± 3.2
	MCHC g/L		347 ± 30		361 ± 30		361 ± 32
	MCHC mmol/L		21.5 ± 2.0		22.4 ± 2.0		22.4 ± 2.0
	RDW %		11.5 ± 3.0		11.5 ± 3.0		10.5 ± 3.0
	PLT × 10 ³ /μL		62 ± 25		248 ± 50		524 ± 70
	MPV fL		7.3 ± 3.0		7.3 ± 3.0		7.7 ± 3.0
	PCT %		0.050 ± 0.050		0.180 ± 0.100		0.410 ± 0.200
	NEU %		26.0 ± 8.0		60.0 ± 8.0		70.5 ± 8.0
	LYM %		65.5 ± 9.0		30.0 ± 7.0		13.5 ± 7.0
	MON %		7.0 ± 7.0		5.0 ± 5.0		2.5 ± 2.5
	EO %		1.0 ± 1.0		4.5 ± 4.5		13.0 ± 13.0
	BA %		0.5 ± 0.5		0.5 ± 0.5		0.5 ± 0.5
	NEU × 10 ³ /μL		0.4 ± 0.2		4.3 ± 0.6		13.8 ± 1.6
	LYM × 10 ³ /μL		1.0 ± 0.2		2.2 ± 0.5		2.6 ± 1.4
	MON × 10 ³ /μL		0.1 ± 0.1		0.4 ± 0.4		0.5 ± 0.5
	EO × 10 ³ /μL		0.1 ± 0.1		0.3 ± 0.3		2.5 ± 2.5
	BA × 10 ³ /μL		0.1 ± 0.1		0.1 ± 0.1		0.1 ± 0.1

WBC Differential alarms may occur with control material. This will not affect the validity of results.

Before using, refer to the instruction sheet for mixing directions.

For technical assistance in the USA and Canada call Technical Service at (800) 523-3395.

All brands and products are trademarks or registered trademarks of their respective companies.

R&D Systems, Inc.
 614 McKinley Place NE
 Minneapolis, MN USA 55413
 AS185-001 Rev. 11/19