



**Günter
Keul GmbH**

CBC-3D

CONTROL

LOT

B0826

ASSAY VALUES AND EXPECTED RANGES

VALEURS CIBLES ET INTERVALLES DE VARIATIONS

RiliBÄK

Wertebblatt



2026-11-05

Applicable Instruments	Parameter	Unit	LOT			LOT			LOT		
			B0826L			B0826N			B0826H		
Keul Hem-o-test 3600	WBC	$\times 10^9/L$	2,00	$\pm$	0,13	7,71	$\pm$	0,50	21,31	$\pm$	1,39
	Lym%	%	59,4	$\pm$	9	30,2	$\pm$	8	12,6	$\pm$	8
	Gran%	%	29,1	$\pm$	9	61,3	$\pm$	8	81,8	$\pm$	8
	Mid%	%	11,5	$\pm$	10	8,5	$\pm$	6	5,6	$\pm$	4
DYMIND DH31 DH33 DH36	Lym#	$\times 10^9/L$	1,16	$\pm$	0,3	2,32	$\pm$	0,7	2,69	$\pm$	2
	Gran#	$\times 10^9/L$	0,57	$\pm$	0,3	4,73	$\pm$	0,7	17,43	$\pm$	2
(Technical File Version A5,0 or higher)	Mid#	$\times 10^9/L$	0,23	$\pm$	0,2	0,66	$\pm$	0,5	1,19	$\pm$	0,5
	RBC	$\times 10^{12}/L$	2,24	$\pm$	0,09	4,67	$\pm$	0,19	5,59	$\pm$	0,22
DH31 Vet DH36 Vet (Technical File Version A1,0 to A4,0)	HGB	g/L	57	$\pm$	2,28	134	$\pm$	5,36	184	$\pm$	7,36
	HCT	%	18,6	$\pm$	0,93	42,8	$\pm$	2,14	57,1	$\pm$	2,86
	MCV	fL	82,9	$\pm$	5	91,6	$\pm$	5	102,2	$\pm$	5
	MCH	pg	25,4	$\pm$	2,5	28,7	$\pm$	2,5	32,9	$\pm$	2,5
	MCHC	g/L	306	$\pm$	30	313	$\pm$	30	322	$\pm$	30
	RDW-CV	%	18,7	$\pm$	3	17,9	$\pm$	3	17,1	$\pm$	3
	RDW-SD	fL	53	$\pm$	8	56	$\pm$	8	58,8	$\pm$	8
	PLT	$\times 10^9/L$	66	$\pm$	9	246	$\pm$	21	483	$\pm$	36
	MPV	fL	8,2	$\pm$	3	8,1	$\pm$	3	8,4	$\pm$	3
	PDW	fL	9,7	$\pm$	3	9,5	$\pm$	3	10,1	$\pm$	3
	PCT	%	0,053	$\pm$	0,05	0,198	$\pm$	0,2	0,405	$\pm$	0,2
	P-LCR	%	15	$\pm$	10	15,4	$\pm$	10	17,1	$\pm$	10
	P-LCC	$\times 10^9/L$	10	$\pm$	10	38	$\pm$	25	82	$\pm$	30
	PDW	/	16,1	$\pm$	3	16,2	$\pm$	3	16,2	$\pm$	3



Bio-technique® - 19 Rue Louis Delourmel

35230 - NOYAL CHATILLON / SEICHE - FRANCE

R&D Systems, Inc. - 614 Mc Kinley Place N.E. - Minneapolis, MN USA 55413

